

Table of Contents

Résumé.....	II
Abstract.....	III
List of Figures.....	VI
List of Tables.....	VIII
List of Symbols and Abbreviations.....	IX
Acknowledgements.....	XI
1. Introduction.....	1
1.1 Background.....	1
1.2 Objectives and Scope of the Project.....	1
2. Literature Review.....	3
2.1 Different Methods of Measuring the Respiratory Rate.....	3
2.2 Oxygen Saturation and Major Analytical Methods.....	5
2.3 Pulse Oximetry Theory Review and Calibration Approach.....	6
2.3.1 Pulse Oximetry Theory Review.....	6
2.3.2 Limitations of the Beer-Lambert Model and Calibration Approach of Commercial Pulse Oximeter.....	8
2.3.3 Pulse Oximetry Signals for Heart Rate Measurement.....	9
2.3.4 Overview of Commercial Pulse Oximeters.....	10
3. Methodology and Design Procedure.....	11
3.1 Methodology.....	11
3.2 Initial Instrument Design for Respiratory Rate Measurement.....	12
3.2.1 Instrument Design Requirements.....	12
3.2.2 Instrument Design Method.....	12
3.2.3 Respiratory Sensor Belt Development.....	13
3.2.4 Microcontroller and Hardware Design.....	13
3.2.5 Software Design.....	16
3.3 Initial Design of the Pulse Oximeter.....	18
3.3.1 Hardware Design and Components Selection.....	19
3.3.2 Algorithm and Software Implementation.....	36
4. Experiments and Results.....	44
4.1 Respiration Rate Monitor System Test.....	44

4.1.1 Material and Method	44
4.1.2 Testing Results	45
4.2 Pulse Oximeter Test	47
4.2.1 Material and Method	47
4.2.2 Testing Results	49
5. Conclusion and Future Work	55
References.....	57
Appendix A Beer's Law and Light Transmittance.....	59
Appendix B Beer-Lambert Model in Pulse Oximetry	62

List of Figures

Figure 2.1 A Two-Electrode Circuit Model.....	4
Figure 2.2 Extinction Spectra of O ₂ Hb and HHb	6
Figure 2.3 Sample Calibration Curve	9
Figure 2.4 Pulsatile Signal while IR or Red light is Transmitted through the Tissue.	10
Figure 3.1 Wireless Instruments Needed for Remote Monitoring of COPD Patients.....	11
Figure 3.2 Real-time Respiration Rate Monitor System Diagram.....	12
Figure 3.3 Conductive Rubber Cord.....	13
Figure 3.4 Conductive Rubber Cord Fixed on an Elastic Band.....	13
Figure 3.5 Arduino Micro.....	14
Figure 3.6 The Sensor Voltage Divider Circuit.....	14
Figure 3.7 Schematic of the Hardware Connection.....	15
Figure 3.8 PCB of the Designed Circuit	16
Figure 3.9 XBee Wireless Module.....	16
Figure 3.10 Program Flow Chart.....	17
Figure 3.11 Pulse Oximeter Design	19
Figure 3.12 Arduino UNO	20
Figure 3.13 Comparison between Transmittance and Reflectance Probes in Pulse Oximetry.....	21
Figure 3.14 Finger Clip Probe along with its Interface Connector Employed in the Design.....	21
Figure 3.15 Schematic of Finger Probe.....	22
Figure 3.16 Schematic of Designed LED Drive Circuit.....	23
Figure 3.17 Red LED Current Control Circuit.....	24
Figure 3.18 IR LED Current Control Circuit.....	25
Figure 3.19 DAC Schematic.....	27
Figure 3.20 Current to Voltage Converter in the Circuit.....	28
Figure 3.21 Bode Plot of the First Active Filter.....	29
Figure 3.22 Output Signal of TIA.....	30
Figure 3.23 Bode Plot of an Ideal Passive Band Pass Filter.....	31
Figure 3.24 Schematic of the Designed Band Pass Filter.....	32
Figure 3.25 Bode Plot of the Designed Band Pass Filter.....	32

Figure 3.26 Signal at the Output of Band Pass Filter.	33
Figure 3.27 Active Low Pass Filter.	34
Figure 3.28 Voltage Divider Circuit.....	35
Figure 3.29 Output of the Last Filter as a Response to the IR LED.	35
Figure 3.30 An IR-light Signal Plot Graph of the Testing Participant.	36
Figure 3.31 A Red-Light Signal Plot Graph of the Testing Participant.....	37
Figure 3.32 Algorithm 1 on IBI, AC and DC values Calculation.	39
Figure 3.33 Algorithm 2 on Heart Rate Calculation.....	40
Figure 3.34 Buffer Shifting for IBI Values.....	41
Figure 3.35 Flow Chart of the Program.....	42
Figure 3.36 Pulse Signals on Oscilloscope during LED Switching.....	43
Figure 4.1 The Designed Respiration Monitor System during the Test.....	44
Figure 4.2 Elapsed time (milliseconds) and the Corresponding Real-time Respiration Rate (breaths per minute) Received from Arduino At Rest State.....	45
Figure 4.3 Elapsed Time (milliseconds) and the Corresponding Real-time Respiration Rate (breaths per minute) Received from Arduino at A Rapid Respiration Rate.....	46
Figure 4.4 Signal Waveforms on Oscilloscope in the Rapid Respiration Test.....	46
Figure 4.5 Comparison of Real-time Test Results between the Prototype and MD300c.....	48
Figure 4.6 Stationary Bike Used in the Test.	49
Figure 4.7 Comparison of Real-time Heart Rate at Rest.....	51
Figure 4.8 Comparison of Real-Time Heart Rate after Prescribed Exercises.....	51
Figure 4.9 Comparison of SpO2 Measurement between the Prototype and MD300c.....	52
Figure 4.10 SpO2 of Participant 1.....	53
Figure 4.11 SpO2 of Participant 2.	54

List of Tables

Table 3.1 Functionality of the ADG884 Input Logic Signal.	24
Table 4.1 Average Results and Percentage of Error for Heart Rate and SpO2.	50

List of Symbols and Abbreviations

COPD	Chronic Obstructive Pulmonary Disease
CHUL	Centre Hospitalier de l'Université Laval
EMFit	Electromechanical Film
ADC	Analog to Digital Converter
DAC	Digital to Analog Converter
PEP	Pyroelectric Polymer
O ₂ Hb	Oxyhemoglobin
HHb	Deoxyhemoglobin
SO ₂	Oxygen Saturation
pH	Power of Hydrogen
PaO ₂	Partial Pressures of Oxygen
SpO ₂	Arterial Oxygen Saturation
IR	Infrared
LED	Light-Emitting Diode
PD	Photodiode
AC	Alternating Current
DC	Direct Current
PCB	Print Circuit Board
PPG	Plethysmography
UART	Universal Asynchronous Receiver/Transmitter
SPDT	Single-pole, Double-throw
EEPROM	Electrically Erasable Programmable Read-only Memory
I ² C	Inter-integrated Circuit
SCL	Serial Clock Line
SDA	Serial Data Line
TIA	Transimpedance Amplifier
BPM	Beats per Minute

IBI	Inter Beat Interval
ISR	Interrupt Service Routine

Rapport-Gratuit.com

Acknowledgements

I would like to thank Prof. **Hung Tien Bui**, my director, for his guidance, encouragement, support, motivation and patience in my study. Special thanks to Prof. **Mario Leone**, my co-director for his patience and guidance. I am particularly grateful for all your help during my studies.

I would like to thank my evaluation committee members for all their guidance through this process; your ideas and feedback have been absolutely invaluable.

I would like to thank all my friends for their support and friendship. I would like to thank my family for their continuing support.

1. Introduction

1.1 Background

Chronic obstructive pulmonary disease (COPD) is a respiratory disorder typically characterized by chronic obstruction of lung airflow that interferes with normal breathing and is not fully reversible [1]. It is a common lung disease in Canada affecting at least 700,000 adults and is now the fourth leading cause of death [2]. People with COPD are likely to present symptoms such as shortness of breath, persistent cough, chest tightness and lack of energy. As COPD sufferers often find the normal daily routines (climbing stairs, shopping, housework, etc.) an uphill struggle, it often leads to a progressively inactive lifestyle. This will cause a vicious cycle that causes health to deteriorate in these patients. Proper rehabilitation training programs have been shown to have beneficial effects on COPD patients [3]. A prescribed exercise program enables COPD patients to increase physical capacity and day to day autonomy. Due to some restricting situations, such as cold weather, limited training space and long distance from training centers, a rehabilitation exercise program designed to be performed at patients' homes may be a good option. Since COPD is a disease that can be deadly, several factors in the training process, such as improper training intensity, duration and frequency may lead the patients to a potentially dangerous situation. Therefore, the patients' safety must be considered during the training program. A real-time monitored training program that assesses key physiological parameters of the patients could be an effective solution. A training expert could then ensure the patients' safety by monitoring the following physiological parameters: respiratory rate, oxygen saturation and heart rate.

1.2 Objectives and Scope of the Project

Recent advances in technology have helped improve the speed and quality of communications media. More particularly, the speed of internet connections has increased dramatically, thus allowing real time transmission of audio and video streams. This has paved the way for applications such as video conferences where individuals from any location can discuss while also establishing visual contact. More recently, certain research groups have proposed it as a tool for medical/health supervision. This would help the population quickly gain access to health

professionals even when they are in remote locations or when travelling to the clinic is not a viable option.

For this reason, researchers at Université du Québec à Chicoutimi (UQAC) and at the Centre Hospitalier de l'Université Laval (CHUL) are developing a hardware/software infrastructure for training of remote patients affected by COPD. With the help of the infrastructure, a health specialist will monitor several COPD patients performing exercises over an internet connection and ensure that the movements are performed correctly and that the exercise intensity remains within prescribed limits. In order to do so, they require a visual connection as well as real-time measurements of physiological signals.

The objective of this master's project was to develop the instruments required to measure the three aforementioned parameters (respiratory rate, heart rate and oxygen saturation). These instruments needed to be small, wireless, low-cost and accurate.

2. Literature Review

In this chapter, we present a detailed analysis of the existing measurement methods related to the desired physiological parameters as well as existing implementations.

2.1 Different Methods of Measuring the Respiratory Rate

Respiratory parameters are important for research in healthcare monitoring. The respiratory rate, which is defined as the number of breaths per minute, is an essential physiological parameter in abnormal breathing diagnosis. Therefore, many systems have been developed for respiratory rate measurements. These systems are mainly based on three kinds of methods: thoracic/abdominal expansion measurement, impedance pneumography and airflow breath measurement.

The first method is based on the fact that the thorax and the abdomen expand and retract with every breath. By measuring the frequency of this expansion/retraction, it is possible to measure the rate at which a person breathes. Over the years, researchers have proposed multiple methods to measure this expansion/retraction. Some researchers have used a mercury filled elastic strain-gauge (Silastic™), whose resistance is sensitive to changes in length. By wrapping this elastic around the torso, the breathing rate can be measured. However, as mercury is a hazardous material, mercury-filled strain gauges are rarely used today. In recent years, wearable sensor-based health monitoring devices, such as BioHarness™ 3 Chest Strap and Hexoskin™ smart shirt have appeared on the market. While they could provide accurate respiratory rate, their relative high costs make them less appropriate for monitoring of a large number of patients. Similarly, Tuomas Reinvuo et al. from the University of Oulu have created a sensor belt with a high-resolution accelerometer and an electromechanical film (EMFit) pressure sensor to measure respiratory rate[4]. Results show that the sensor belt is a portable and accurate tool for respiratory rate measurement. However, the system is only suitable for the immobile testing.

The second technique, known as impedance pneumography, is based on the fact that air causes a change in electrical impedance in the lungs. Therefore, by measuring the rate of change in impedance in the lungs, it would be possible to determine the breathing rate. To use this approach, either two or four electrodes are attached to the chest. The impedance in the thorax can

be divided into two components: a relatively constant baseline impedance R_B and a varying respirative impedance ΔR [5]. During inspiration, the increase in gas volume of the chest and the expansion of the thorax cause ΔR to increase[6]. The relationship between the total impedance($\Delta R+R_B$) and the volume of air in the lungs is approximately linear[7]. If electrical current was applied to the electrodes during the breathing process, the change in impedance (ΔR) will result in a change in voltage (ΔV). To perform breathing rate measurement, a possible approach is to inject a high-frequency AC current into the body through the electrodes. This ac signal acts as a carrier signal that is amplitude-modulated by the change in voltage caused by respiration (ΔV). After demodulation and denoising with a low-pass filter, the output can then be digitized by an analog to digital converter (ADC). The respiratory rate then could be calculated. Figure 2.1 shows a typical two-electrode circuit model [7]. In this figure, R_p are the defibrillator protection resistances within the cables that are in series with the electrode impedances ($Z_{\text{Electrode}}$) which consist of a 51-k Ω resistor and 47-nF capacitor in parallel [3].




Figure 2.1 A Two-Electrode Circuit Model

The third type of approach used in respiration measurement consists of measuring the air exhaled. For this approach, it is customary to use a face mask to either measure the air displacement or the change in temperature due to respiration. Dodds et al. proposed the use of a pyroelectric polymer (PEP) film as a transducer for respiratory rate measuring[8]. The PEP film is a material that generates a voltage when its temperature or pressure is changed (pyroelectric effect). In their proposed system, the material is mounted to the internal surface of a standard oxygen face mask. The voltage variation, caused by the change in temperature during respiration, is sent to an ADC after signal processing. Results show that the PEP film is a low-cost and

reliable transducer for respiratory rate monitoring. The major disadvantage with this approach is the use of a face mask which may cause discomfort in many participants and the need daily maintenance for proper hygiene.

2.2 Oxygen Saturation and Major Analytical Methods

Oxygen transport is important in sustaining human life. Hemoglobin is a protein that is capable of carrying and releasing molecular oxygen (O_2) in erythrocytes (red blood cells). Hemoglobins can be divided into oxyhemoglobin (O_2Hb or HbO_2) and deoxyhemoglobin (HHb or Hb), according to whether they are fully saturated with oxygen or not. Most of the hemoglobin in human blood is either in the form of O_2Hb or b . For new born babies with certain respiratory problems and patients under anesthesia, their blood oxygenation needs to be monitored since they are unable to breathe on their own. It is important to determine whether they are absorbing the necessary amount of oxygen supplied by ventilators. Similarly, for patients with certain respiratory problems, such as COPD, oxygen absorption could fall to dangerously low levels. Therefore, the assessment of oxygen saturation (SO_2)— a parameter that indicates the ratio of the oxyhemoglobin over both oxy and deoxyhemoglobins, provides a valuable indicator in respiration monitoring.

The following empirical equation is used to determine hemoglobin oxygen saturation, using measured hemoglobin concentrations:

$$SO_2 = \frac{cO_2Hb}{cO_2Hb+cHHb} \quad (2.1)$$

A normal range of SO_2 for healthy adults, expressed as a percentage, is typically 94%–98%. For patients who are not achieving the critical blood oxygen saturation level of 90%, proper treatments may be required [9].

There are three major analytical methods used to measure oxygen saturation: arterial blood gas analyzers, CO-oximetry and pulse oximetry [10]. The arterial blood gas analyzers calculate estimated oxygen saturation in a blood sample based on empirical equations using power of hydrogen (pH) and the partial pressures of oxygen (PaO₂) values. The CO-oximetry measures the absorption of hemoglobin derivatives from blood samples using multiple wavelengths of

light. While they are precise, the above two methods require blood samples and do not support real-time monitoring and therefore, are not suitable for our project.

2.3 Pulse Oximetry Theory Review and Calibration Approach.

2.3.1 Pulse Oximetry Theory Review

Pulse oximetry is a noninvasive method of continuously assessing arterial oxygen saturation (SpO₂ or SaO₂) and the pulse rate of a patient. It is widely prevalent in health care and is often regarded as the fifth vital sign after body temperature, pulse rate, respiration rate and blood pressure[11]. By measuring light absorption at a well-perfused body area (finger, ear lobe or nasal alar) with two different wavelengths (red and infrared), it assesses the SpO₂ and the pulse rate.

The general principle of pulse oximeters is explained as follows. According to the optical properties of *O₂Hb* and *HHb* (Figure 2.2), the absorption of *O₂Hb* and *HHb* at red and near-infrared (IR) light are significantly different. *HHb* allows more IR light (940 nm) to pass and absorbs more red light (660nm), while *O₂Hb* behaves in the opposite way. This is in accordance with the fact that well-oxygenated blood with higher concentration of *O₂Hb* often appears redder to the eye than poorly-oxygenated blood.




Figure 2.2 Extinction Spectra of *O₂Hb* and *HHb* [12].

It is also important to note that the body tissues do not typically attenuate the optical signals at these two wavelengths whereas yellow, green, blue, and far- IR light signals are significantly absorbed by vascular tissues and water[13].

Based on the above difference in light absorption properties of O_2Hb and HHb , a pulse oximeter typically emits two wavelengths of light, red light at 660 nm and near- IR light at 940 nm from two light-emitting diodes(LEDs). These lights will then transmit through the finger, ear lobe or other body tissues and will be detected by a photodiode or photo detector (PD). The LEDs and PD are commonly integrated together in a device called a probe in a pulse oximeter. The light signal received on the PD is converted into an electrical signal. thus, the relative absorption of red and IR light can be quantified.

If we assume the transmission of light through the arterial blood is influenced only by the relative concentrations of O_2Hb and HHb and their absorption coefficients at the two measurement wavelengths, the light intensity will decrease logarithmically with path length according to the Beer–Lambert law[14] (see Appendix A). With these principles, the ratio of the intensity of light transmitted at two different wavelengths could be expressed as follow:

$$R = \log_{10} (I_1) / \log_{10} (I_2) \quad (2.2)$$

In Equation (2.2), I_1 is the intensity of light with wavelength λ_1 (660 nm) and I_2 is the intensity of light with wavelength λ_2 (940nm).

Once the absorbance coefficients of O_2Hb and HHb at the two wavelengths are known, the oxygen saturation can be obtained using the following equation:

$$SpO_2 = (\alpha r_2 * R - \alpha r_1) / [(\alpha r_2 - \alpha o_2) * R - (\alpha r_1 - \alpha o_1)]. \quad (2.3)$$

Where:

- αr_1 is the extinction coefficient of HHb at λ_1
- αr_2 is the extinction coefficient of HHb at λ_2
- αo_1 is the absorption coefficient of O_2Hb at λ_1
- αo_2 is the absorption coefficient of O_2Hb at λ_2

- R is the ratio obtained from Equation (2.2)
- SpO₂ is the arterial oxygen saturation

This SpO₂ calculation principle is based on the Beer–Lambert model (see Appendix B).

2.3.2 Limitations of the Beer-Lambert Model and Calibration Approach of Commercial Pulse Oximeter

In the Beer-Lambert model, the arterial blood is treated as a homogeneous absorbing medium. In real situations, the absorbance of light is not simply proportional to the concentration of hemoglobin or to the optical path length but are also dependent on scattering and multiple scattering. This phenomenon can occur when light is refracted by a similar-sized object to the wavelength of the light, as in the case of red/IR light having the same size wavelength as red blood cells (approximately 7 μm in diameter). Scattering causes the deviation of a light beam from its initial direction and therefore, highly increases light absorbance. In addition, light that is scattered once will likely be scattered again by cells and therefore multiple scattering occurs[14].

Therefore, the Beer-Lambert model can't be applied directly to the SpO₂ calculation in pulse oximetry. As the process of mathematically modeling the problem of light scattering for different conditions is very complex, most commercial pulse oximeters use calibration curves from empirical formulas to determine the SpO₂. This method provides SpO₂ values that are accurate enough for clinical use.

Instead of calculating the standard R value using Equation (2.2), an approximation of the R value is typically employed in pulse oximetry as the error is negligible (below 0.02%)[15].

$$R = (AC_{\text{red}} / DC_{\text{red}}) / (AC_{\text{IR}} / DC_{\text{IR}}) \quad (2.4)$$

In the equation, AC_{red} and AC_{IR} are the absorbance of pulsatile “alternating current” component of the red and IR light. DC_{red} and DC_{IR} is the absorbance of non-pulsatile “direct current” (DC) component of the red and IR light.

Therefore, R can be obtained from the AC/DC ratio of red and IR light signals. In the pulse oximeter algorithm, the continuously calculated R value is used together with a look-up table which is made up of empirical formulas to determine the SpO₂. As an important part of the pulse oximeter system, the look-up tables are usually based on calibration curves generated empirically

from measuring the R values of many healthy subjects with SpO2 levels altered from 100% to approximately 70%[16]. The SpO2 of the healthy subjects is usually measured by a very accurate method such as co-oximeter. Figure 2.3 shows a sample calibration curve.




Figure 2.3 Sample Calibration Curve.

Since generating a custom calibration curve is both time consuming and expensive, Equation (2.5) is often suggested in the literature[14].

$$\text{SpO}_2 = 110 - 25 * R \quad (2.5)$$

This equation is a linear approximation of an empirical calibration curve established by measurements of a large group of healthy volunteers with SpO2 values generally greater than 70%. Therefore, it could be used to compute the SpO2 in our design.

2.3.3 Pulse Oximetry Signals for Heart Rate Measurement

In pulse oximetry, when the IR or red light is transmitted through the tissue, it produces a pulsatile signal as shown in Figure 2.4 [17]. The signal varies with time in relation to the heart beat. Therefore, the heart rate frequency of the individual can be extracted from the waveform.



Figure 2.4 Pulsatile Signal while IR or Red light is Transmitted through the Tissue.

2.3.4 Overview of Commercial Pulse Oximeters

With recent technological advancements and increasing interest in physiological measurements, low-cost pulse oximeters have become widely available on the market. Wireless portable pulse oximeters have also emerged, such as the health Air PO3 and the Massimo MAS-9809. In these products, the measured SpO₂ and heart rate are transferred to Apple iOS or Android devices with their customer application programs via Bluetooth LE connectivity[18, 19]. These devices can be very useful in typical situations. However, in this project, the wireless pulse oximeter is used together with the wireless respiration belt to monitor the patient's physiological parameters. Hence, to synchronize the data, the pulse oximeter needs to be compatible with the already designed wireless respiration belt. Unfortunately, the wireless communication standard used by the commercial pulse oximeters on the market (Bluetooth LE) could not interface with the XBee wireless module that was used in the designed respiration monitoring system[20]. Meanwhile, designing a new pulse oximeter in the project could also pave the way for more integrated measurements systems. For example, researchers have demonstrated the possibility of detecting respiratory signals in plethysmograms obtained from pulse oximeters[21]. With a modified pulse oximeter, it would be possible to measure the three-mentioned key physiological parameters using a single piece of equipment. Therefore, it was deemed important to design a pulse oximeter for this project.

3. Methodology and Design Procedure

3.1 Methodology

For the purpose of the master's project, two wireless instruments were developed to monitor 3 key physiological parameters, respiratory rate, oxygen saturation and heart rate. This helps ensure the safety of the COPD patients during their exercise program. Figure 3.1 shows the principle of our wireless system when used in the project.




Figure 3.1 Wireless Instruments Needed for Remote Monitoring of COPD Patients.



3.2 Initial Instrument Design for Respiratory Rate Measurement

3.2.1 Instrument Design Requirements

For measuring of the respiratory rate of COPD patients, the following basic requirements should be met. First, as the COPD patients will use this system during their indoor exercise program, the system should be portable and should be able to acquire and transfer real-time results without disturbing the patients' exercise. Second, since the patients' respiratory rates can vary due to their exercise intensity, the testing system should be accurate for the real-time monitoring. A low-cost system is preferred with wide measuring range and quick response.

3.2.2 Instrument Design Method

As introduced in the literature review, the expansion and contraction of thorax is an indicator for the respiratory rate measurement. According to the design requirements, our design is based on this principle. By measuring the time elapsed between two consecutive thorax expansions, the participant's respiratory rate is acquired.

The respiration rate monitoring system consists a respiratory sensor belt, a micro-controller and a wireless communication unit. The wearable system keeps measuring the patients' respiratory rates and sends them to a computer (Figure 3.2[22]).




Figure 3.2 Real-Time Respiration Rate Monitor System Diagram.

3.2.3 Respiratory Sensor Belt Development

The sensor belt is designed to acquire the raw signal of breathing events from the COPD patients. In our design, a conductive rubber cord is selected for the sensor belt due to its functionality and low cost (Figure 3.3). The cord is made of carbon-black impregnated rubber and has an electrical resistance that increases gradually when stretched. With the conductive rubber cord strapped around the chest, the thoracic or abdominal contraction and expansion during breathing could be converted to electrical signals. The rubber cord is fixed on an elastic band and used as a sensor belt for the respiratory measurement (Figure 3.4).



Figure 3.3 Conductive Rubber Cord.



Figure 3.4 Conductive Rubber Cord Fixed on an Elastic Band.

3.2.4 Microcontroller and Hardware Design

Due to its simplicity, small size and its relatively good computing power, Arduino Micro is selected as the microcontroller board for the project. It is the smallest board of the Arduino family and is suitable for the portability demand of the project (Figure 3.5).




Figure 3.5 Arduino Micro.

To convert the change in electrical resistance to a change in electrical signal, the rubber cord is connected to a resistor to form a voltage divider (Figure 3.6). The output of this voltage divider is fed into an analog input of the Arduino.




Figure 3.6 The Sensor Voltage Divider Circuit.

The output of the voltage divider is given by Equation (3.1).

$$V_{out} = V_{CC} * \frac{R}{R_{strap} + R} \quad (3.1)$$

In the equation, V_{CC} is the supply voltage, R_{strap} is the resistance of sensor belt and R is a constant resistor. This voltage is read into the Arduino's analog input pin A5.

In the design, as the breathing signal produced by the sensor belt is very small, two amplifiers are required for the signal processing. The INA126 was chosen as the amplifier of choice since it is a precision instrumentation amplifier that is accurate, low noise and provides differential-signal acquisition[23]. According to the datasheet, the output of the amplifier, given by:

$$VO = (Vin^+ - Vin^-) G. \quad (3.2)$$

$$\text{Where, } G = 5 + 80k\Omega / R_G$$

The Vin^+ pins of the amplifiers in Figure 3.7 are connected to the Arduino digital output pins to give a reference value. As the input analog signal changes with the expansion/retraction of the chest, the reference value is set as the average input to the Vin^- , so that the breathing signal is amplified properly and output of amplifier is zero when there is no breathing.




Figure 3.7 Schematic of the Hardware Connection.

For the first amplifier, R2 and C1 constitute a first order passive low pass filter, where the cut-off frequency (f_c) is given by using the Equation (3.3).

$$f_c = 1/(2\pi C * R). \quad (3.3)$$

This circuit filters the high frequency noise from the output of the Arduino's digital output pin and feeds the resulting signal into the input of the INA126. Similarly, R3 and C2 fulfill the same function for the second amplifier. The outputs of the two amplifiers are connected to the Arduino's analog input pins A1 and A3 for breathing rate calculation. Figure 3.7 shows the general schematic diagram of the hardware connection. Figure 3.8 is the PCB (print circuit board) of the hardware. The Arduino micro is connected to the testing board during the testing.



Figure 3.8 PCB of the Designed Circuit.

A pair of XBee wireless communication modules (part number XB24-Z7WIT-004) from Digi International were selected in the project to provide wireless connectivity between the Arduino board and the computer. The modules use the IEEE 802.15.4 networking protocol for fast point-to-multipoint networking with low power consumption. In the wireless respiration rate monitor system, the transmitter module is connected to the PCB of the testing board and the receiver module is connected to a computer via an adapter board.



Figure 3.9 XBee Wireless Module.

3.2.5 Software Design

The software of the system is programmed in C with the Arduino Software (IDE). The running process of the program is shown below.




Figure 3.10 Program Flow Chart.

After the initialization of variables, the program executes the *setup()*. In this part, it keeps reading the analog signal of the pin A5 and A1 for 6 seconds each to get the average breathing signal values from the sensor belt. Then, these values are written to V_{in}^+ of the corresponding amplifiers as reference values. Afterwards, the program waits for the first breathing and memorizes the first breathing event system time T1 using the *millis()* function. The program

waits for the sensor belt signal to go low for a certain time T before detecting the next breathing signal T_2 . This adjustable time T helps reduce false detections by setting the minimum time interval between two valid consecutive respiration events. When the second breathing event occurs, system time T_2 is measured and the program calculates the elapsed time $T_2 - T_1$. Thus, the elapsed time is the time interval between the two consecutive respirations. Finally, the real-time respiratory rate R (breaths per minute) could be calculated with Equation (3.4).

$$R = 6000 / (T_2 - T_1). \quad (3.4)$$

After calculation, the R value is wirelessly transferred to a computer through the serial communication port. The program runs in a loop and keeps detecting the breathing event and sending the data during the testing.

As noises are usually introduced by the sensor belt and the hardware itself, we used an existing library to smooth out the analog output signals by averaging consecutive output readings.

3.3 Initial Design of the Pulse Oximeter

In this section, the design process of the pulse oximeter is presented. Before introducing the implementation details, a functional block diagram of the design is presented (Figure 3.11).

The optical system in the diagram emits red and IR lights alternatively under the control of the LED drive circuit and microcontroller. Thus, a plethysmography (PPG) signal of the test subject is produced on the side of PD. The signal is then amplified and filtered before being sent to the 10-bit ADC module of the microcontroller. At the microcontroller side, the analog PPG signal (0 to 5volts) is converted to digital numbers (0 to 1023) with the built-in ADC. With the proper algorithm, the R value in Equation (2.4) and the heart rate are calculated. The desired data from the microcontroller is transferred to the computer using the wireless module via its universal asynchronous receiver/transmitter port (UART). The detailed process is presented in the following sections.




Figure 3.11 Pulse Oximeter Design

3.3.1 Hardware Design and Components Selection

3.3.1.1 Microcontroller Selection

The microcontroller is at the heart of the system. It controls the entire system and deals with the signals coming from the sensors. Considering the hardware requirements, the Arduino UNO board was chosen for this application (Figure 3.12). The board is designed with an Atmel328P processor and combines many features such as a 10-bit ADC, internal and external interrupts, 14 digital I/O pins and 6 analog input pins, which are all necessary for the design.



Figure 3.12 Arduino UNO


3.3.1.2 Optical Probe Selection

As mentioned above, LEDs and PD are required to send and receive optical signals. Depending on the relative position of LEDs and PD, there are two methods of sending the light to the measuring tissues: transmissive method and reflective method.

The transmissive method is a traditional method applied in most of the commercial pulse oximeters. For this method, the LEDs and PD are placed right opposite to each other with the measuring site between them. The red or IR light emitted from the LEDs will travel through the measured sites. Then, the residual light will reach the PD after the absorption. The typical measuring sites for the transmittance probe are finger, ear lobe, nasal alar, toe for adults or the foot or palm for infant. The reason for targeting these body tissues is that there is a much higher vascular density than other body areas[24]. A clip-like probe is often used for this method to hold the measured tissues.

The reflective method was first introduced by Brinkman and Zijlstra in 1949[25]. For this method, the LEDs and PD are placed on the same side of the measuring site with the PD receiving the reflected light signal from different depths underneath the skin. This method is becoming more and more popular in recent years due to its flexibility of measuring sites. A variety of body sites can be measured such as forehead, wrist, ankle, ear, etc.

Since tissue is highly forward scattering, the relative number of photons detected in reflective mode is low. This will cause a relatively poor PPG signal. Therefore, for this method, sensors should be well designed for better acquisition of the radial reflectance profile[26].



(A) Transmittance Probe Example Used in this Project. (B) Reflectance Probe Example.

Figure 3.13 Comparison between Transmittance and Reflectance Probes in Pulse Oximetry[27, 28].

Both two optical arrangement methods could be used in the oximeter design. For the traditional transmissive method, as the amount of light passing through the tissue is concentrated at the PD, it has better light intensity at the PD compared to the reflective method. In our project, a finger clip transmittance reusable probe (model number: Nellcor DS-100A) was selected for its simplicity and high PPG signal strength (Figure 3.14).




Figure 3.14 Finger Clip Probe along with its Interface Connector Employed in the Design.

There are two small high-intensity red and IR LEDs on one side of the probe and a silicon PD on the opposite side of the LEDs. A spring is used in the probe to adjust the probe to the finger sizes



of testing subjects. The probe is connected to the outer circuit through a 7-pin sub-D connector (DB9). The schematic of the probe is shown in Figure 3.15.




Figure 3.15 Schematic of Finger Probe.

3.3.1.3 LED Drive Circuit

In the optical system of pulse oximeter, the red and IR LEDs are switched to emit light passing through the testing finger. For different testing subjects, the residual light intensity at the PD may differ from each other. This may be caused by the difference of finger thickness, arterial blood volume and skin absorption. Thus, the light intensity needs to be adjusted for different participants. For the optical system, we need to control both the illumination time and the intensity of the red and IR light through the microcontroller. Figure 3.16 shows the schematic of designed LED drive circuit. As shown in the probe schematic, the two LEDs are connected back to back and only the red or the IR light can be turned on at a time. Based on the LED configuration, an analog switch was used to control the LED in the design. For the light intensity, it was adjusted by controlling the current through the LEDs. A digital to analog converter (DAC) was used in the circuit to implement the current control.




Figure 3.16 Schematic of Designed LED Drive Circuit.

In the circuit, the ADG884BRMZ was selected as the analog switch. The component is a low voltage device containing two independently selectable single-pole, double-throw (SPDT) switches. Each switch on the device works between 1.8V to 5.5V with a high current handling capacity (400 mA continuous current). Considering all the features above, the ADG884 is a suitable choice for the application.

In the schematic (Figure 3.16), COM1 and COM2 are connected to the anode and cathode of the LEDs. NC1 and NC2 are connected to the collectors of the 2N3904 NPN type bipolar junction transistor (BJT). Pin IN1 and IN2 of ADG884 are connected to the Arduino digital I/O pin to receive a high or low digital command.

Table 3.1 shows the functionality of ADG884 with two kinds of input logic signals. According to the table, when Arduino digital output D4(IN1) is 1 and D7(IN2) is 0, NC1 and NO2 are off and NC2 and NO1 are on.

Table 3.1 Functionality of the ADG884 Input Logic Signal.

Functionality of the ADG884			
Input Logic Signal	NC1 and NO2	NC2 and NO1	LED status
IN1=0& IN2=1	ON	OFF	Red off & IR on
IN1=1& IN2=0	OFF	ON	Red on & IR off

In other words, the red LED anode is connected to the 3.3V Vcc and cathode is connected to the collector of Q2. The equivalent circuit diagram is shown in Figure 3.17.




Figure 3.17 Red LED Current Control Circuit.

Similarly, when the Arduino digital output D4 (IN1) is 0 and D7 (IN2) is 1, the equivalent circuit for the IR light becomes that of Figure 3.18.




Figure 3.18 IR LED Current Control Circuit.

Therefore, the switch of two LEDs is controlled by the digital signal from the Arduino digital I/O pins.

For a LED, the current flowing from the anode to the cathode is defined as forward current with the typical values ranging from 2 to 50 mA. The forward current determines the light intensity of the LED. Hence, we can control the light intensity of the LEDs via the forward current. The potential drop across the p-n junction of the diode (forward voltage) is often between 0.9 and 2.5 V.

Considering the circuit in Figure 3.9, the transistor Q1 acts as a controlled current sink, where the current flowing in the base is determined by the base voltage of Q1, the resistor R1 and R3. For the transistor with a gain of β , the forward current I_C is β times of the base current I_B . When the transistor is turned on, there is a voltage drop across the base emitter junction which is about 0.7V.

If the DAC output is V_{DAC} and the base voltage of the transistor is V_B , the current flow the base of transistor I_B and the emitter current I_E is expressed as follow.

$$I_B = (V_{DAC} - V_B)/R1 \quad (3.5)$$

$$I_E = (V_B - 0.7)/R3 \quad (3.6)$$

Since $I_C = \beta * I_B$ and $I_E = I_B + I_C$, after derivation with Equation (3.5) and (3.6), we have the LED forward current in Equation (3.7).

$$I_C = \beta * (V_{DAC} - 0.7) / (R1 + R3 * (\beta + 1)). \quad (3.7)$$

Since V_{DAC} is tightly controlled by the microcontroller, the gain of the transistor β can be considered constant and $R1$ and $R3$ are fixed values, the forward current I_C is essentially determined by the microcontroller. In the circuit, the anode of the LED is connected to the 3.3V power supply to provide enough forward voltage drop.

For the DAC, the MCP4728 from Microchip Technology was selected in the design. The device is a 12-bit DAC with nonvolatile memory (EEPROM) and four output channels. It uses a two-wire I²C (Inter-Integrated Circuit) serial interface to communicate with the Arduino and works as a slave device when connected to the I²C bus line. The configuration such as input codes, device configuration bits, and address bits are programmable to its nonvolatile memory by using I²C serial interface commands. The nonvolatile memory makes it possible to hold the DAC settings during power-off time, allowing the outputs to be available immediately after power-up. In the design, a 0.1 μ F bypass capacitor together with an additional 10 μ F capacitor in parallel was connected to the power supply to attenuate the high-frequency noise from the application board. The 5.1k Ω pull-up resistors for SCL and SDA are suitable for Standard (100 kHz) and Fast (400 kHz) modes (Figure 3.19).




Figure 3.19 DAC Schematic.

Two output channels of the device are connected to the base terminals of Q1 and Q2. A 1 nF capacitor is used to stabilize the output voltage. The channel output voltage is associated with its configuration bit settings and DAC input code. It is calculated with the equation below.

$$V_{\text{out}} = V_{\text{DD}} * D_n / 4095. \quad (3.8)$$

In this equation, D_n is the DAC input value (0 to 4095) that is set in the Arduino program and V_{DD} is the reference voltage. Therefore, the red and IR light forward currents are controlled by the Arduino microcontroller.

3.3.1.4 Signal Conditioning

The PD in the probe generates a current when receiving the light that travels through the finger. Since the light received by the PD also includes ambient light, the current generated by the PD will also be mixed with noise. This current is a small signal and that is generally a few hundreds of nanoamps according to our tests. To get a clean and proper PPG signal in order to determine the oxygen saturation and heart rate, the raw signal needs to be amplified and filtered before being sent to the Arduino.

The PD current needs to be converted to voltage before it is filtered. A transimpedance amplifier (TIA) is used for this purpose. The precision operational amplifier LT1677 from Linear Technology was selected in the project for its rail-to-rail input and output, low input bias current and low noise features. Figure 3.20 presents the implemented circuit with LT1677.




Figure 3.20 Current to Voltage Converter in the Circuit.

In the circuit, the input resistance of the amplifier is so high that all the current from the PD must go through the feedback resistor R11. The gain of the TIA is therefore determined by R11. If the current from the PD cathode is I , the output voltage will be given by Equation (3.9). Since the current is very small, a 4.7 Mohm resistor was selected for the required gain.

$$V_{out} = I \cdot R_{11}. \quad (3.9)$$

A capacitor C6 is placed in parallel with the transimpedance resistor R11. The TIA, R11 and C6 together constitute a first order active low pass filter, where the cut-off frequency (f_c) is given by using the Equation (3.10).

$$f_c = 1 / (2\pi C \cdot R). \quad (3.10)$$

As this is the first stage of the low pass filter, only the high frequency noises will be removed. Thus, a 125Hz cut-off frequency was chosen and this was accomplished using a 270 pF capacitor after calculation with Equation (3.10). The Bode plot of the filter is presented in Figure 3.21.




Figure 3.21 Bode Plot of the First Active Filter[29].

At low frequencies, the input signal is passed directly to the output with a gain of the TIA until it reaches the cut-off frequency f_c at which point the gain is $1/\sqrt{2}$ (-3dB) of the maximum gain. After this point, the response of the circuit is attenuated with slope of -20dB/Decade.

The non-inverting input of the LT1677 is connected to a voltage from the voltage divider. Therefore, the initial amplitude level of the TIA output is shifted. This step converts the positive and negative output signal of TIA into a positive-only range signal for further conditioning. Finally, the signal is shifted to a range between 0 and 5V when sent to the Arduino ADC. With Equation (3.11), it is possible to calculate the output of the voltage divider network. The level shift voltage is set to 0.35V:

$$V_{out} = R_9 * V_{CC} / (R_9 + R_{10}). \quad (3.11)$$

Figure 3.22 shows the output the of the TIA with the previously described configuration.




Figure 3.22 Output Signal of TIA.

Figure 3.22 shows that the average level of the output also changes with time. This is often attributed to breathing which changes in intrathoracic pressure throughout the respiratory cycle[30]. The figure also shows that there are high frequency noises. This means both low and high frequency noises are present in the system, and that a band pass filter should be used.

The band pass filter in our design consists of a low pass filter, a high pass filter and a buffer between them. The buffer has a very high input impedance and a small output impedance which provides electrical isolation between the two filter circuits to facilitate the design process. The voltage gain of the filter is calculated using Equation (3.12).

$$\text{Gain} = 20 \log(V_{\text{out}}/V_{\text{in}}). \quad (3.12)$$

Equation (3.10) can be used to calculate the cut-off frequency of the passive band pass filter. The characteristic of an ideal band pass filter is presented by the Bode Plot in Figure 3.23.




Figure 3.23 Bode Plot of an Ideal Passive Band Pass Filter[31].

As shown in Figure 3.23, at low frequencies, the gain of the filter changes with a slope of +20dB/Decade until the frequency reaches the cut-off frequency of the low pass filter f_L . The gain continues to be at its maximum value until it reaches the cut-off frequency of the high pass filter f_H , then it is attenuated with a slope of -20dB/Decade.

To calculate the cut-off frequency of the band pass filter, the characteristics of the signal of interest (the heart rate) are analyzed. A normal resting heart rate for adults ranges from 60 to 100 beats per minute (BPM) whereas a well-trained athlete might have a resting heart rate closer to 40 BPM[32]. In the current project, heart rates between 40 and 180 BPM (0.7 Hz and 3Hz) will be considered. Therefore, the pass band of the filter will be between 0.7 Hz and 3Hz. The cut-off frequency of each stage of the passive band pass filter was calculated using Equation (3.10). The designed band pass filter is shown in Figure 3.24.






Figure 3.24 Schematic of the Designed Band Pass Filter.

Figure 3.25 shows the frequency response curve of the designed band pass filter.




Figure 3.25 Bode Plot of the Designed Band Pass Filter.

An output example signal from oscilloscope is shown in Figure 3.26.




Figure 3.26 Signal at the Output of Band Pass Filter.

As the figure shows, the low frequency noise was attenuated compared to Figure 3.22 and there were still some high frequency noises in the system. Since the band pass filter is a passive filter the gain of signal is also smaller compared to the output of the TIA.

In the filter (Figure 3.24), the output of the buffer is connected to the Arduino ADC to read the DC level which is an important parameter in the oxygen saturation calculation. The reason for reading the DC value at this point is that the DC signal is removed after the high pass filter.

Since the passive band pass filter can't remove all the noise, a first order active low-pass filter was introduced (Figure 3.27). LT1677 operational amplifier was also used in this filter design.




Figure 3.27 Active Low Pass Filter.

The gain can be calculated as follows:

$$\frac{V_{in}-0}{R15} = \frac{0-V_{out}}{\frac{1}{\frac{1}{R14}+j\omega c9}} \quad (3.13)$$

After derivation, we have equation below.

$$\frac{V_{out}}{V_{in}} = \frac{H_0 \omega_0}{S+\omega_0} \quad \text{with} \quad \left| \frac{V_{out}}{V_{in}} \right| = |H_0| \frac{\omega_0}{\sqrt{\omega_0^2+\omega^2}} \quad (3.14)$$

In this equation, H_0 is $-R14/R15$ and ω_0 is $1/(R14*C9)$. For this low pass filter, the cutoff frequency is calculated with Equation (3.10) where $R14$ is used for calculation. At high frequencies ($\omega \gg \omega_0$) the capacitor acts as a short circuit and the gain of the amplifier drops to zero. At very low frequencies ($\omega \ll \omega_0$) the capacitor is an open and the gain of the circuit is H_0 .

It is important to note that the DC voltage of $V_{cc}/2$ (V_{cc} is 3.3V) is supplied to $R13$ and the non-inverting input of the op-amp (Figure 3.24 & Figure 3.27). The purpose of this input voltage is to superimpose the AC signal to a DC level equal to $V_{cc}/2$. As stated in the previous section, to

interface with the Arduino ADC, the amplitude range of the AC voltage should be 0 to 3.3V. To generate this voltage, a voltage divider circuit with a LT1677 amplifier (Figure 3.28) was introduced. The operational amplifier buffer in the circuit was also used to eliminate the interaction with the adjacent circuit.




Figure 3.28 Voltage Divider Circuit

With all the described signal conditioning, the resulting signal was suitable for interfacing with the Arduino ADC (Figure 3.29).



Figure 3.29 Output of the Last Filter as a Response to the IR LED.

During the hardware testing, the circuit is implemented on a breadboard since it was simpler to make adjustments as required. Integrating the components and the microcontroller onto a PCB would make the device more portable and easier to use in the further research.

3.3.2 Algorithm and Software Implementation.

After the signal conditioning, the amplified red and IR PPG signals were ready for analysis. This section presents how the analog signal was processed in the Arduino program to calculate the SpO₂ and heart rate.

3.3.2.1 Signal Feature Analysis

Before explaining the detailed algorithm, an analysis of the filtered signal is presented.

The ADC of the Arduino Uno can map the input voltages (0 to 5V) into integer values ranging from 0 to 1023 every 100 microseconds. Thus, the red or IR output voltages (output in Figure 3.27) can be converted into digital values. Figure 3.30 and Figure 3.31 present the plotted red and IR digital values in MATLAB. These values were obtained by testing a participant with the designed hardware.




Figure 3.30 An IR-light Signal Plot Graph of the Testing Participant.




Figure 3.31 A Red-Light Signal Plot Graph of the Testing Participant.

In the two above figures, the x-axis and the y-axis are unitless. The y axis is the digitized voltage with a range of 0 to 1023. The x-axis is the sample number. The sample frequency of the Arduino ADC is 100KHz.

As the IR light penetrates the body deeper than red light, it generates a larger peak to peak value. It is interesting to note that, when the signal goes down to the base level, there is a small dip (also called dicrotic notch)[33]. It is caused by the closure of the aortic valve and it needs to be considered in the algorithm to avoid calculation errors.

The time interval between the two consecutive heartbeats is also known as the inter beat interval (IBI). A sample IBI is shown in Figure 3.31 and it will be explained in detail in the following chapter. The trough and peak values correspond to the points where the wave presents the smallest and largest ADC values, respectively. The *Thresh level*, as indicated in the figure, corresponds to half of the sum of peak and trough values.

3.3.2.2 Software Design

The heart rate can be calculated with the IBI. To perform this task, the moment at which a heartbeat occurs can be determined using several methods. Some researchers consider the point at which the signal reaches 25% or 50% of maximum height of the PPG signal[34]. In this project, the IBI is obtained by measuring the time interval between two adjacent wave points where the fast-rising wave reaches 50% of its maximum value. Consider Figure 3.31, for example. The time interval between A and B is a sample IBI.

To calculate SpO₂, as stated in Section 2.3, it is calculated using Equation (2.5). To obtain the R value in this equation, the red and IR LEDs are switched on and off to measure their relative AC and DC values (see Equation (2.4)). The AC value of the red or IR signal is the peak to trough value (see Figure 3.31). The DC value is the average value of the signal.

To measure the IBI, the Arduino needs to have a regular sampling rate. For that purpose, the program uses a timer module and sets an interrupt every 2ms (500 Hz sampling rate). When the interrupt is triggered, the program will pause its current activity and execute the interrupt service routine (ISR).

When the program begins, it sets up different variables including the DAC output parameters. Afterwards, the red LED is switched on and finally, the timer interrupt is started.

The IBI, AC and DC values are the key factors used to calculate the heart rate and SpO₂. The pseudo-code of Algorithm 1 in Figure 3.32 explains how these values are determined during a pulse signal.

Algorithm 1

Calculate the IBI, AC and DC values during a pulse signal. The function is triggered every 2ms

```
1 Read filter output value A at Pin A0;
2 current Time = last current Time + 2ms ;
3 interval N = Current Time - the last heart beat time
4 if ( the pulse is going down )
5 {
6     find the trough value
7 }
8 if (the value A bigger than the middle of the pulse)
9     find peak value
10 if (the time interval N >300ms)
11 {
12     if (the pulse passes the middle of the wave)
13         calculate the time interval since last beat
14         Update last heart beat time
15 }
16     if (the wave is going down)
17 {
18         AC value = peak value- trough value
19         update the Thresh value
20         update the peak value and trough value
21         Read DC output value B;
22 }
```

Figure 3.32 Algorithm 1 on IBI, AC and DC values Calculation.

In the algorithm, the interrupt is executed every 2ms. Thus, the current time is updated at the beginning accordingly (line 2). To determine the pulse is going down, two conditions need to be satisfied (line 4). First, the value A needs to be less than the threshold value (450 by default and updates every pulse). Second, interval N is bigger than 3/5 of the previous IBI. This step is to make sure the pulse wave has passed the dicrotic notch (Figure 3.31). The trough (minimum) value of the wave is found here by comparing value A (line 6). When the interval N is bigger than 300ms, the program begins to detect the pulse events (line 10). This step is the same as a low pass filter where the passing frequency is calculated by $1/(300\text{ms})$. To attenuate signals with frequencies higher than the respiration frequencies, 300ms is set as the default value in the program. If 3/5 of the last IBI has passed and the value A is bigger than the threshold value, the pulse is detected. Meanwhile, IBI is obtained by subtracting last beat time from the current time N. When the value A is smaller than the threshold value, the wave is going down. The peak and trough values found through this algorithm help determine the AC value of the signal. The DC value is obtained by reading the DC output from the circuit

With the algorithm above, the IBI time, the AC and DC values can be calculated for a single beat. Instantaneous values, however, are prone to noise. In addition, the IBI tends to vary from one heartbeat to the next and therefore, an average value would be more interesting for our application. Therefore, we implemented a moving average filter as shown in Figure 3.33.

Algorithm 2

buffer shifting algorithm to get an averaged IBI and calculate BPM

```

1 IBI = Sample Time – last heart beat Time;
2 if (the first beat is true) {
3   wait second beat
4 }
5 if (the second beat is true) {
6   for item 0 to 9
7     set IBI_ARRAY [item] = IBI
8   end for
9 }
10 else {
11   for index 0 to 8
12     shift the array item as Figure 3.23 indicates
13     calculate the sum of IBI_ARRAY
14   end for
15   IBI_ARRAY [9] = new tested IBI
16   sum IBI = sum IBI + IBI_ARRAY [9]
17   Average of the array : AVERAGE is SUM_IBI / 10
18   Heart rate BPM is 60000 / AVERAGE
19 }

```

Figure 3.33 Algorithm 2 on Heart Rate Calculation.

In Figure 3.33, since the IBI is calculated by subtracting the previous beat time, the IBI from the first beat is discarded. The array of IBI is initialized by giving the new IBI value to each of its item at the second beat (line 6 and 7). Beginning from the third one, the IBI values are processed by the moving average method. The heart rate is calculated from the average of the IBI.

As the program runs, the new IBI value are placed in the position 9 of the array with the older values being shifted to the right. The value at position 0 is discarded when new data arrive. Figure 3.34 illustrates this algorithm. The output of the heart rate is determined by the average of last nine pulse waves. Through experiments, it was found that 9 heart beats were the minimum value that would provide good results.




Figure 3.34 Buffer Shifting for IBI Values.

The flowchart in Figure 3.35 describes the program. To calculate the SpO₂, one of the LEDs remains ON for 8 heart beats while the other LED is OFF. The roles are then inverted for the following 8 heart beats. The values from the 16 heart beats are then used to calculate SpO₂.

For most of pulse oximeters on the market, the LED is switched rapidly between red and IR (about several milliseconds). In the prototype, the LED is switched every 8 heart beats. The benefit of this new approach is that it reduces the influence of light switching noise. Meanwhile, as the value of SpO₂ is calculated every 8 heart beats, the result is already an averaged representative value.




Figure 3.35 Flow Chart of the Program.

During the experiment, it was noted that, when the LED changes, the first wave is not usable for AC calculation. This can be seen at point A in Figure 3.36. Thus, the first pulse is discarded when calculating the AC value.




Figure 3.36 Pulse Signals on Oscilloscope during LED Switching.

When the program begins, it initializes the variables and parameters in the program. The DAC and interrupt register parameters are set up at this point. It then proceeds to the interrupt and the main loop. The interrupt function is triggered and calculates the heart rate and SpO₂ values (see Figure 3.32 and Figure 3.33). The main loop begins to send values to the serial port when the interrupt function gives it a signal. In the interrupt function, as the values arrive, they are used to calculate the heart rate and SpO₂. The LED is switched when the detected pulse number is equal to 8.

4. Experiments and Results

After the instruments were designed, a series of experiments were carried out to evaluate the performance of the designed prototypes. For the respiration rate monitoring system, the measurements were compared to visual observations. For the pulse oximeter prototype, the measurements were compared to the results of a reference commercial pulse oximeter under the same testing conditions. In this section, we present the test methods and the results obtained.

4.1 Respiration Rate Monitor System Test.

4.1.1 Material and Method

After respiration rate monitor system was designed, a series of experiments were carried out to test of the designed prototype.

Two healthy (one female and one male), non-smoking students aged 25 and 32 participated the evaluation. In order to participate in the study, the students had to be free from respiratory diseases, obesity, cardiovascular problems. The ambient air temperature in the laboratory is about 23°C during the whole test.




Figure 4.1 The Designed Respiration Monitor System during the Test.

Two measurements were performed under different conditions for the analysis. The first measurement evaluated the participants at rest and the latter one was performed when the participants had a higher breathing rate. Since breathing rate could be self-controlled, the participants were instructed to breathe rapidly to get a higher breathing rate. An oscilloscope was used to monitor the analog signals sent to the Arduino during the test. The real-time breathing time and breathing rate were sent to a computer via XBee wireless transmission modules.

4.1.2 Testing Results

During the test, the respiration rate monitor system presented high accuracy and every respiration event was detected by the instrument. For each respiration event, the time interval between the two consecutive respirations (elapsed time) and the real-time respiratory rate were sent to a computer through wireless transmission.

Figure 4.2 and Figure 4.3 present the elapsed time (in milliseconds) and the corresponding real-time respiration rate (breaths per minute) received from Arduino at the two different respiration events of a participant.




Figure 4.2 Elapsed time (milliseconds) and the Corresponding Real-time Respiration Rate (breaths per minute) Received from Arduino At Rest State.




Figure 4.3 Elapsed Time (milliseconds) and the Corresponding Real-time Respiration Rate (breaths per minute) Received from Arduino at A Rapid Respiration Rate.

During the test, the respiratory events of the participants were monitored visually and showed high accordance with the prototype measuring results. Meanwhile, on the oscilloscope, the waveform of each respiration was monitored during the test. Figure 4.4 presents oscilloscope waveform corresponding to the last four respiratory events of Figure 4.3. We can find the estimated elapsed times from the waveform are consistent with the received data in Figure 4.3.




Figure 4.4 Signal Waveforms on Oscilloscope in the Rapid Respiration Test.

According to the testing results, the respiration rate monitor system presented high accuracy in real-time respiration detection and is ready to be used for training COPD patients in remote locations.

4.2 Pulse Oximeter Test

4.2.1 Material and Method

A group of five healthy, non-smoking students (2 females and 3 males) participated in the validation of the prototype. They ranged in age between 22 and 32 years (25.6 ± 3.7 years). In order to participate in the study, the students had to be free from respiratory diseases, obesity, cardiovascular problems or severe contact allergies to latex or other materials in pulse oximetry sensors. The ambient air temperature in the laboratory is about 23°C during the whole test.

To provide reference to validate the prototype, an oximeter is needed for the tests. As introduced in Section 2.2, the arterial blood gas analyzers and CO-oximetry are precise to measure the oxygen saturation. Since they require blood samples and do not support real-time monitoring, they are not suitable to provide reference. A commercial pulse oximeter from Choice Technology (MD300c) was used to provide real-time reference in the tests. An oscilloscope was used to monitor the analog input signals to the Arduino at the same time.

At the beginning of the tests, the DAC output parameters were adjusted for each participant to obtain a proper signal level. During the test, the probe of the prototype was attached to the middle finger. The reference oximeter was attached to the index finger for comparison (Figure 4.5). For each participant, the test results were measured continuously over a time-span of 3 minutes. The heart rate and calculated R values from the designed prototype were printed to the screen while the values of the MD300c were recorded at the same time (Figure 4.5). For the MD300c, since it had no digital or analog output for the results, readings were recorded by hand throughout the test.




Figure 4.5 Comparison of Real-time Test Results between the Prototype and MD300c.

Two measurements were performed under different conditions. The first measurement evaluated the participants at rest and the latter one was performed after a moderate-intensity aerobic physical activity on a stationary bike. Before the exercise, the bike was adjusted to a proper preset resistance level and each participant took a 5-minute warmup on the bike to prepare for the test (Figure 4.6). Since the pulse oximeter is very sensitive to motion, the participants were asked to limit their movements during the test phase.



Figure 4.6 Stationary Bike Used in the Test.

Average heart rate and SpO₂ measured using the MD300c and the prototype were calculated after the experiment. The percentage error was also selected for the result analysis using Equation (4.1).

$$\text{Percentage Error} = \left| \frac{\text{Reference value} - \text{Prototype value}}{\text{Reference value}} \right| * 100 \quad (4.1)$$

Simple linear regression method was used for modeling the results relationship between the MD300c and the prototype.

4.2.2 Testing Results

The average results and the percentage of error for heart rate and SpO₂ of all participants are listed in Table 4.1.

Table 4.1 Average Results and Percentage of Error for Heart Rate and SpO2.

Reference Pulse Oximeter		Prototype Pulse Oximeter		Percentage Error	
SpO2 (%)	Heart Rate	SpO2 (%)	Heart Rate	SpO2	Heart Rate
98.25	96.93	96.30	97.41	1.96	0.50

For the heart rate tests, the test results of the prototype showed high consistency and accuracy compared to the outputs of the reference MD300c during the entire test.

The calculated heart rate percentage error between the prototype and the MD300c was 0.5% (Table 4.1) which means that the designed prototype showed high accuracy in the heart rate tests. The test results were also analyzed by linear regression. Figures 4.7 and 4.8 show the comparison of heart rate data measured by the prototype and the reference MD300C at rest and after exercise. The solid line represents the linear regression line of the test results. The reference dashed line represents the linear line on the assumption that the prototype and the reference MD300C results are equal. The relative R^2 are separately 0.95 and 0.93 under the two conditions, indicating that the data measured by the prototype device correlates very well with the reference device.




Figure 4.7 Comparison of Real-time Heart Rate at Rest.




Figure 4.8 Comparison of Real-Time Heart Rate after Prescribed Exercises.

For the SpO₂ tests, as stated in the literature review, Equation (2.5) is selected as a linear approximation to the empirical calibration curve. The measured results were rounded to the nearest decimal. This is conducive to analyzing the trends of SpO₂ changes.

The calculated SpO₂ percentage error between the prototype and the MD300c was 1.96% (Table 4.1). During the test phase, we found the SpO₂ results of the reference MD300c stayed relatively stable between 97% and 99%. At the same time, the results fluctuated between 93% and 99% for our prototype.

Figure 4.9 shows the comparison of the oxygen saturation data. The solid line represents the linear regression line. The reference dashed line represents the assumption that the prototype and the reference MD300C results are equal. The R² value is equal to 0.56, indicating that the data measured by the prototype device does not correlate very well with the reference device. We believe that this may due to the empirical calibration equation and the MD300c measurement accuracy, as the measurement accuracy of MD300C is ±2% in the test measurement range.[35]




Figure 4.9 Comparison of SpO₂ Measurement between the Prototype and MD300c.

When the test results of each participant are plotted individually in the time domain, they all showed some common features. Here are the results of two participants (Figure 4.10 and Figure 4.11). The results are obtained under the following conditions. At the beginning, the participants are at rest. After receiving an instruction, the participants began to breathe deeply to induce more oxygen into the bloodstream. After several seconds, the SpO2 values from our prototype began increasing earlier than the reference MD300c. This suggests that the designed prototype can provide a more sensitive output than the reference device which is important for the project.




Figure 4.10 SpO2 of Participant 1.




Figure 4.11 SpO2 of Participant 2.

5. Conclusion and Future Work

The aim of this research was to design a low cost wireless system to measure three physiological parameters: respiratory rate, oxygen saturation and heart rate for training COPD patients remotely.

For the respiratory rate measurement, a respiratory sensor belt was designed to measure the respiratory rate. A conductive rubber cord was selected to construct the sensor belt due to its functionality and low cost. Filters and amplifiers were designed for signal conditioning and an algorithm was designed on a microcontroller to measure the respiration rate. The final output was sent to a computer using Xbee wireless modules. It was evaluated with experiments and demonstrated high accuracy for respiratory detection.

For the oxygen saturation and heart rate monitoring, a pulse oximeter was designed and built to measure these parameters at the fingertip. A group of volunteers participated in the test program in the laboratory. A commercial pulse oximeter was used as a reference for performance comparison. Two evaluation metrics, linear regression and percentage error, are used for the result analysis.

The heart rate test results of the prototype demonstrated high consistency and accuracy comparing to the results of commercial reference MD300c during the whole test.

The SpO₂ test results are within the normal SpO₂ level (94% to 99%) even though correlation measures were low. This can be explained by the use of an empirical equation and accuracy errors from the commercial pulse oximeter. It is interesting to note that the designed oximeter shows good sensitivity when the participants are breathing and therefore, allows for the quick detection of health problems relating to oxygen saturation. The prototype device measurement accuracy could be improved if calibrated with CO-oximetry in the further research.

The output of the prototype is sent using the same wireless transfer method (via Xbee wireless modules) as presented in Section 3.2.4.

The designed system meets the required objectives by successfully measuring the respiration rate, the pulse rate and oxygen saturation, and could wirelessly transmit the data.

While this research project provided a proof of concept, further research is required in order to provide a more thorough scientific validation of the proposed devices.

References

1. WHO. *Definition of COPD*. 2015 [cited 2015 December 15]; Available from: <http://www.who.int/respiratory/copd/definition/en/>.
2. O'Donnell, D.E., et al., *Canadian Thoracic Society recommendations for management of chronic obstructive pulmonary disease–2008 update–highlights for primary care*. Canadian respiratory journal: journal of the Canadian Thoracic Society, 2008. **15**(Suppl A): p. 1A.
3. Strijbos, J.H., et al., *A comparison between an outpatient hospital-based pulmonary rehabilitation program and a home-care pulmonary rehabilitation program in patients with COPD: a follow-up of 18 months*. CHEST Journal, 1996. **109**(2): p. 366-372.
4. Reinvoio, T., et al. *Measurement of respiratory rate with high-resolution accelerometer and EMFit pressure sensor*. in *Sensors Applications Symposium, 2006. Proceedings of the 2006 IEEE*. 2006. IEEE.
5. Grenvik, A., et al., *Impedance pneumography: Comparison between chest impedance changes and respiratory volumes in II healthy volunteers*. Chest Journal, 1972. **62**(4): p. 439-443.
6. Kelkar, S.P., N.D. Khambete, and S.S. Agashe, *Development of movement artefacts free breathing monitor*. J. Instrum. Soc. India, 2004. **38**: p. 34-43.
7. Gupta, A.K. *Respiration Rate Measurement Based on Impedance Pneumography*. 2011 [cited 2015 December 12]; Available from: <http://www.ti.com/lit/an/sbaa181/sbaa181.pdf>.
8. Dodds, D., J. Purdy, and C. Moulton, *The PEP transducer: a new way of measuring respiratory rate in the non-intubated patient*. Journal of Accident & Emergency Medicine, 1999. **16**(1): p. 26-28.
9. EasyOxygen. *Oxygen Saturation levels and what do they mean?* [cited 2015 12-20]; Available from: <http://www.easyoxygen.com.au/oxygen-saturation-levels-and-what-do-they-mean/>.
10. Haymond, S. *Oxygen Saturation A Guide to Laboratory Assessment*. 2006 [cited 2015 12-21]; CLINICAL LABORATORY NEWS]. Available from: <http://www.optimedical.com/pdf/articles/oxygen-saturation-laboratory-assessment.pdf>.
11. Mower, W.R., et al., *Pulse oximetry as a fifth pediatric vital sign*. Pediatrics, 1997. **99**(5): p. 681-686.
12. Ekkekakis, P., *Illuminating the black box: investigating prefrontal cortical hemodynamics during exercise with near-infrared spectroscopy*. Journal of sport & exercise psychology, 2009. **31**(4): p. 505.
13. Chan, E.D., M.M. Chan, and M.M. Chan, *Pulse oximetry: Understanding its basic principles facilitates appreciation of its limitations*. Respiratory Medicine, 2013. **107**(6): p. 789-799.
14. Webster, J.G., *Design of pulse oximeters*. 1997: CRC Press.
15. Tavakoli Dastjerdi, M., *An analog VLSI front end for pulse oximetry*. 2006, Massachusetts Institute of Technology.
16. Schnapp, L.M. and N.H. Cohen, *Pulse oximetry: uses and abuses*. Chest, 1990. **98**(5): p. 1244-1250.
17. Townsend, N., *Medical Electronics. Michaelmas Term, 2001*.
18. Masimo. *MightySat™ Rx Fingertip Pulse Oximeter*. [cited 2017 June 04]; Available from: <http://www.masimo.com/home/signal-extraction-pulse-oximetry/masimo-set-monitors/mightysatrx/>.
19. iHealth. *iHealth Air Wireless Pulse Oximeter*. [cited 2017 June 04]; Available from: <https://ihealthlabs.com/fitness-devices/wireless-pulse-oximeter/>.
20. Lee, J.-S., Y.-W. Su, and C.-C. Shen. *A comparative study of wireless protocols: Bluetooth, UWB, ZigBee, and Wi-Fi*. in *Industrial Electronics Society, 2007. IECON 2007. 33rd Annual Conference of the IEEE*. 2007. Ieee.
21. Leonard, P., et al., *Standard pulse oximeters can be used to monitor respiratory rate*. Emergency Medicine Journal : EMJ, 2003. **20**(6): p. 524-525.

22. FaCt.Inc. *Physiological Status Monitoring* [cited 2015 12-21]; Available from: <http://www.fact-canada.com/BioHarness/Zephyr-Bioharness.html>.
23. Instruments, T. *INAx126 Datasheet MicroPower Instrumentation Amplifier* 2015 [cited 2015 12-21]; Available from: <http://www.ti.com/lit/ds/symlink/ina126.pdf>.
24. Mannheim, P.D., *The light-tissue interaction of pulse oximetry*. *Anesthesia & Analgesia*, 2007. **105**(6): p. S10-S17.
25. Brinkman, R. and W. Zylstra, *Determination and continuous registration of the percentage oxygen saturation in clinical conditions*. *Archivum chirurgicum Neerlandicum*, 1949. **1**(3): p. 177.
26. Li, K. and S. Warren, *A wireless reflectance pulse oximeter with digital baseline control for unfiltered photoplethysmograms*. *IEEE transactions on biomedical circuits and systems*, 2012. **6**(3): p. 269-278.
27. Electronics, M. *Reflective optical sensor well suitable for Pulse Oximeter and Heart rate monitor*. 2014 [cited 2017 2-21]; NJL5501R photo incation]. Available from: <https://rutronik-tec.com/njl5501r/>.
28. Sang-Soo Oak, P.A., *Design Peripheral of Oxygen Saturation (SpO2)* 2017.
29. *Frequency Response of a 1st-order Low Pass Filte*. 2017 [cited 2017 2-21]; Available from: http://www.electronics-tutorials.ws/filter/filter_2.html.
30. Leonard, P., et al., *Standard pulse oximeters can be used to monitor respiratory rate*. *Emerg Med J*, 2003. **20**(6): p. 524-5.
31. *Frequency Response of a 2nd Order Band Pass Filter*. 2017; Available from: http://www.electronics-tutorials.ws/filter/filter_4.html.
32. Edward R. Laskowski, M.D. *What's a normal resting heart rate?* 2015 [cited 2017 2-22]; Available from: <http://www.mayoclinic.org/healthy-lifestyle/fitness/expert-answers/heart-rate/faq-20057979>.
33. *Arterial Pressure Waveforms*. [cited 2017 Mars 10]; Available from: <http://www.dynapulse.com/educator/webcurriculum/chapter%203/arterial%20pressure%20waveforms.htm>.
34. Smith, R.P., et al., *Pulse transit time: an appraisal of potential clinical applications*. *Thorax*, 1999. **54**(5): p. 452-457.
35. *MD300C1 Fingertip pulse oximeter*. [cited 2017 Mars 11]; Available from: http://www.amperorblog.com/doc-lib/MD300C1%20Technical%20Specification%20VER_1.0.pdf.
36. Glaros, K.N., *Low-power pulse oximetry and transimpedance amplifiers*. 2011: Imperial College London (University of London).
37. *Pulse Oximeter Spring 2011 calibration section*. [cited 2017 June 04]; Available from: <https://sites.google.com/site/pulseoximeterspring2011/calibration>.

Appendix A Beer's Law and Light Transmittance

Beer-Lambert's law (also referred to as Beer-Lambert's or Bouguer's law) describes the attenuation of (monochromatic) light when it travels through a uniform light absorbing substance. If a monochromatic incident light of intensity I_0 enters the medium, one part of this light is transmitted through the medium and another part is absorbed. The intensity I of light traveling through the medium decreases exponentially with distance.

$$I = I_0 e^{-\varepsilon(\lambda)cd} \quad (\text{A.1})$$

Where:

- I is the intensity of transmitted light
- I_0 is the intensity of incident light
- $\varepsilon(\lambda)$ is the extinction coefficient or absorptivity of the absorbing substance at a specific wavelength
- c is the concentration of the absorbing substance
- d the optical path length through the medium

In Beer's law, the sum of transmitted and absorbed light equals the incident light. The physical processes of light reflection at the medium surface and the light scattering in the medium is not accounted. Figure A.1 is an illustration of Equation (A.1) in diagram.




Figure A.1 A Diagram Illustrates the Absorption of Light Traveling through a Uniform Medium following Beer's law[14].

From Equation (A.1), the ratio of transmitted light I to the incident light I_0 is defined as the transmittance (T), when light traveling through a medium with an absorbing substance.

$$T = \frac{I}{I_0} = e^{-\epsilon(\lambda) c d}. \quad (\text{A.2})$$

The unscattered absorbance (A) is defined as the negative natural logarithm of the transmittance of light.

$$A = -\ln T = \epsilon(\lambda) c d \quad (\text{A.3})$$

When there is more than one light absorbing substance in the medium, the Beer's law is still valid. Each absorber will contribute its part to the total absorbance. For a medium with n individual absorbing substances, the resulting total light absorbance A_t equals the sum of n independent absorbances.

$$A_t = \epsilon_1(\lambda) c_1 d_1 + \epsilon_2(\lambda) c_2 d_2 + \dots + \epsilon_n(\lambda) c_n d_n = \sum_{i=1}^n \epsilon_i(\lambda) c_i d_i. \quad (\text{A.4})$$

In the equation, for each substance i , $\epsilon_i(\lambda)$ is the extinction coefficient, c_i is concentration and d_i is the optical path length through the absorbing substance.

Therefore, for a homogeneous medium of multi absorbing substances, when there are different wavelengths emitted to the substances, Beer's law can help to determine the light absorbance.

Appendix B Beer-Lambert Model in Pulse Oximetry

In pulse oximetry, the oxygen saturation of arterial blood is measured by calculating the light absorbance of living tissue at two different wavelengths (red light at 660nm and IR light at 940nm) and using the arterial pulsation to differentiate between absorbance of arterial blood and other absorbers.

According to the Beer's law in Appendix A, if we assume the optical path length d for the oxygenated and reduced hemoglobin is the same, with Equation (A.4) and (2.1), the total light absorbance A_t that contains only oxygenated and reduced hemoglobin as absorbing substances is derived as Equation (B.1).

$$A_t = [\epsilon_{\text{HbO}_2}(\lambda)SO_2 + \epsilon_{\text{Hb}}(\lambda)(1 - SO_2)](c_{\text{Hb}} + c_{\text{HbO}_2})d. \quad (\text{B.1})$$

In this equation, A_t is expressed for known concentrations of hemoglobin in terms of arterial oxygen saturation as a fraction, the extinction coefficients of hemoglobin and the length of the optical path.

During a cardiac cycle, the arterial blood volume increases during the systole and decreases during the diastole. On the contrary, the blood volume in the veins and capillaries as well as the volumes of skin, fat, bone, etc., remain relatively constant[13]. Thus, the optical path length d of Equation (B.1) in the arteries is changed with cardiac cycle and keeps relatively constant in other tissues. That is to say, in pulse oximetry, the light absorption in arterial blood of the measuring tissues fluctuates along with cardiac cycles and keeps relatively stable in other areas of the measuring tissues. Therefore, when the LEDs light from the pulse oximeter passes through the measuring tissues and finally reaches the PD, a combined signal with pulsatile “alternating current” (AC) component and non-pulsatile “direct current” (DC) component is generated. The time varying signal of transmitted light generated at PD is referred to as the plethysmography (PPG) signal which is the main signal for SpO2 and pulse rate calculation.

Figure B.1 shows the relation of light absorbed and transmitted in a body tissue as a function of time in pulse oximetry. Usually, the alternating part of the light absorbed by the body tissue does not exceed 1% to 2% of the constant absorbance of the DC components[14].




Figure B.1 PPG Signal of Light Absorption of Body Tissue[36].

The intensity of transmitted light reaching the PD can be expressed as a function of the diameter d of the arteries. It changes from d_{\min} to d_{\max} in a cardiac cycle. Therefore, if we substitute d with $d_{\min} + \Delta d$, after derivation, the transmitted light at PD is expressed as a function of I_{\max} and Δd .

$$I = I_{\max} e^{-[\varepsilon_{\text{Hb}}(\lambda)c_{\text{Hb}} + \varepsilon_{\text{HbO}_2}(\lambda)c_{\text{HbO}_2}]\Delta d} \quad (\text{B.2})$$

The ratio of the red and IR absorbance of the light signal is introduced to calculate the SpO₂.

$$R = \frac{A_{t,R}}{A_{t,IR}} = \frac{\ln(I_{\min,R} / I_{\max,R})}{\ln(I_{\min,IR} / I_{\max,IR})} \quad (\text{B.3})$$

Assuming the red light optical path lengths equals the IR light optical path lengths, using Equation (B.1) and (B.2), we have Equation (B.4).

$$R = \frac{\varepsilon_{\text{Hb}}(\lambda_R) + [\varepsilon_{\text{HbO}_2}(\lambda_R) - \varepsilon_{\text{Hb}}(\lambda_R)]SpO_2}{\varepsilon_{\text{Hb}}(\lambda_{IR}) + [\varepsilon_{\text{HbO}_2}(\lambda_{IR}) - \varepsilon_{\text{Hb}}(\lambda_{IR})]SpO_2} \quad (\text{B.4})$$

Rewrite Equation (B.4), we get the SpO2 function calculated by ratio R.

$$SpO2 = \frac{\varepsilon_{Hb}(\lambda_R) - \varepsilon_{Hb}(\lambda_{IR})R}{\varepsilon_{Hb}(\lambda_R) - \varepsilon_{HbO_2}(\lambda_R) + [\varepsilon_{HbO_2}(\lambda_{IR}) - \varepsilon_{Hb}(\lambda_{IR})]R} \times 100\% \quad (B.5)$$

Where:

- $\varepsilon_{Hb}(\lambda_R)$ is the extinction coefficient of *HHb* at red light
- $\varepsilon_{Hb}(\lambda_{IR})$ is the extinction coefficient of *HHb* at IR light
- $\varepsilon_{HbO_2}(\lambda_R)$ is the extinction coefficient of *O₂Hb* at red light
- $\varepsilon_{HbO_2}(\lambda_{IR})$ is the extinction coefficient of *O₂Hb* at IR light
- *R* is the ratio obtained from Equation (B.3)
- *SpO2* is the arterial oxygen saturation

Therefore, the oxygen saturation in arterial blood is derived theoretically by using Equation (B.3) and (B.5) with Beer's law.

For the Beer-Lambert model, the incident light passing through human tissue is simply split into absorbed light and transmitted light. While in the real situation, some parts of the light are reflected and others are scattered. Thus, the SpO2 results calculated with Beer-Lambert model may differ greatly with the actual SpO2 level. Figure B.2 plots the theoretical calibration curve of the Beer-Lambert model with an empirical calibration curve for comparison. The Beer-Lambert model tends to give erroneous estimates of the true value of oxygen saturation. In the figure, if the SpO2 calculated by empirical calibration curve is 100%, with the same R value, the Beer-Lambert model result is about 93%. This difference between the theoretical curve and the empirical curve is mainly due to light scattering effects.




Figure B.2 Calibration Curves for Pulse Oximeters[37].